



## Laboratory Alert

**Date:** August 16, 2016

**Effective Date:** 8/8/16 – 8/15/16

**Title:** Nitrofurantoin Sensitivity Report Omitted

**Test:** Urine Culture & Sensitivity

**Explanation of change:** Due to a software upgrade problem, Nitrofurantoin was omitted from Urine Culture Sensitivity reports from 8/8/16 until 8/15/16. This is specific to Urine Cultures only and the problem has been resolved. Corrected reports were issued 8/15/16. Please contact the Cayuga Medical Laboratory at 274-4474 if you need further assistance.

  
\_\_\_\_\_  
Date 08-16-16

Dr. Elizabeth Plocharczyk, Laboratory Assistant Medical Director

  
\_\_\_\_\_  
Date 8-16-16

Toni Burger, Laboratory Administrative Director