



Laboratory Technical Bulletin

Date: 4/9/2021

Effective Date: 1/14/21

Title: Hemoglobin A1C Testing 12/20/2020-1/6/2021

Test: A1C

Explanation: Due to an instrument malfunction resulting in a calibration error, patient Hemoglobin A1C test results reported between 12/10/20 and 1/6/21 may have been high by 1 to 3 %. The instrument has been optimized. The issue was resolved on 1/6/21. Hemoglobin A1C testing resumed at CMC Main Lab on 1/6/21 and continues without issues.

Date

04/09/21

Dr. Elizabeth Plocharczyk, Laboratory Medical Director

Date

4/9/21

Patricia Lindquist, Assistant Technical Director